



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
NATIONAL EXPOSURE RESEARCH LABORATORY
RESEARCH TRIANGLE PARK, NC 27711

March 19, 2015

OFFICE OF
RESEARCH AND DEVELOPMENT

To whom it may concern,

Subject: US EPA Approval for airpointer

EPA herewith confirms that the airpointer modules for measuring ambient concentrations of specified air pollutants have been formally designated as either a Federal Reference Method (FRM) or Federal Equivalent Method (FEM) in accordance with Title 40, Part 53 of the Code of Federal Regulations (40 CFR Part 53). Currently approved airpointer analyzers are listed below along with their approved designation numbers.

| | |
|--|--|
| airpointer® module 801-001000 for SO ₂ | Automated Equivalent Method: EQSA-0495-100 |
| airpointer® module 801-004000 for O ₃ | Automated Equivalent Method: EQOA-0992-087 |
| airpointer® module 801-003000 for CO | Automated Reference Method: RFCA-1093-093 |
| airpointer® module 801-002000 for NO/NO ₂ | Automated Reference Method: RFNA-1194-099 |

The listing of airpointer under the related Teledyne methods does not imply any limitations but only states the use of assemblies of this manufacturer by recordum.

Further details can be found in the US EPA list of designated instruments:

<http://www.epa.gov/ttn/amtic/files/ambient/criteria/reference-equivalent-methods-list.pdf>

Best regards,

Robert W. Vanderpool, Ph.D.

A handwritten signature in blue ink, appearing to read "Robert W. Vanderpool".

Manager, Reference and Equivalent Methods Program
U.S. Environmental Protection Agency
Office of Research and Development
National Exposure Research Laboratory
Human Exposure and Atmospheric Sciences Division
Process Modeling Research Branch
109 TW Alexander Drive, Room D475, MD-D205-03
Research Triangle Park, NC 27711
919-541-7877 (voice), 919-541-4848 (fax)
Vanderpool.Robert@epa.gov